

EXPERIMENT: DETERMINATION OF SODA ASH**INTRODUCTION**

The measurement of soda ash in industrial samples is important. In this experiment, an HCl solution is prepared and standardized with dried sodium carbonate. The standardized HCl is then used to analyze a dried soda ash sample.

In the following procedures both indicator and pH meter methods (potentiometric titration) will be used.

PROCEDURE:

Indicator Method (note that the indicators are added only for reference purposes; the important data are the pH readings)

Prepare 500 mL of a 0.1M HCl (conc. HCl stock is approx. 12.1 M) solution by diluting the appropriate quantity of conc. HCl with water. Standardize against sodium carbonate as follows.

Weigh accurately two samples of 0.2g pure, dried sodium carbonate (by difference) into a 250ml Erlenmeyer flask. Add 50mL distilled water and 1-2 drops phenolphthalein indicator. Titrate slowly (0.5mL/sec) until the pink color disappears. This is a slight excess for half the volume of acid required for total neutralization. Add two drops of methyl orange indicator, continue titrating until the color changes from straw to pale orange to pink.

Now weight accurately two 0.3 to 0.4g samples of the dried unknown soda ash and titrate with the standardized acid in duplicate with indicators as above.

At the start of the experiment on the unknown, the student should run a quick, approximate titration using the indicators to establish the approximate location of the end points and the color changes to be expected. With this information, the student should be able to determine the HCl required for the particular end-point, and then carefully add the remaining 10% of the HCl slowly so that an accurate endpoint is obtained.

pH Meter Titration

Weigh accurately 0.2g pure, dried sodium carbonate (by difference) into a 400mL beaker provided with a magnetic stirring bar. Add 50mL water, dissolve. (It is not necessary to add an indicator in this titration since you are to use the pH values to determine the equivalence point more accurately than an indicator can). Insert a glass and a calomel electrode into the solution, avoiding contact with the stirring bar. Add titrant and take readings at 1.0 mL intervals. Record where you think the end points are but don't forget that you still have to do a Gran Plot to ascertain the equivalence points. Exceed the final equivalent volume with an additional 10ml of titrant. Do this titration rapidly, because the pH will tend to drift as CO₂ escapes from the solution.

Now weigh accurately 0.3 to 0.4 g of the dried soda ash unknown. Titrate, following the above procedure using the pH meter. For each pH meter titration, a *mandatory Gran plot* is to be plotted for both the first and second equivalence points. Calculate your results based on the total volume needed to reach the second equivalence point* as if the soda

ash unknown were sodium carbonate and contained no sodium bicarbonate. Report the results as % sodium carbonate (Na_2CO_3) and as % sodium oxide (Na_2O). Note that Na_2O can be thought of as the residue remaining after CO_2 was removed from the sodium bicarbonate.

Possible Sources of Determinate Errors.

If you compare the pH at the first end point with the pH at which you observed the phenolphthalein end point, you may find that the indicator color changes, not at the end point, but by as much as 3% to 5% of the volume of titrant after the endpoint. (For example, the equivalence point may occur at pH 8.4, whereas the red color may disappear at pH 8.0). This has some interesting consequences. One is that, with indicator, the ratio of the volumes of acid at the second to the first end point is not exactly '2' but about 1.9 (check your indicator data in the textbook to see whether the discrepancy occurs and if it agrees with this discussion). Another major difference between the two end points is that of precision. Refer to your data obtained from the pH meter titrations. The main factor influencing precision will be the ratio of changes in pH to changes in volume of added titrant at the first and second end points. To see this, examine the pH vs. V (HCl) graph. At each end point measure the V (HCl) at $[\text{pH}(1)+0.5]$ and at $[\text{pH}(1)-0.5]$, where pH (1) is the pH at the first end point (1), and similarly at the second end point (2). From this, calculate V (HCl) needed to change the pH by ± 0.5 for the first and second end points.

*Calculate your results based on a **Gran Plot** (see Section 11-5 of your textbook for further discussion of the theory). The Gran Plot is based on the following equation:

$$V_a 10^{+\text{pH}} = \left(\frac{1}{K_a} \frac{\gamma_B}{\gamma_{\text{BH}^+}} \right) (V_e - V_a)$$

where V_a = is the volume of strong acid added, pH = is the pH meter reading, K_a = acid dissociation constant of the acid form of the base, BH^+ , B is the weak base being titrated, and V_e = equivalence point. Note that the above equation resembles that for a straight line: $y = mx + b$, for regions not very near the equivalence point. The variables are ($V_a \cdot 10^{+\text{pH}}$) for the ordinate (y-axis) and (V_a) for the abscissa (x-axis). It is not necessary to know the actual values of the other parameters, K_a , γ_B , γ_{BH^+} and V_e . By extrapolating the straight line portion of the graph to obtain the x-intercept, one can determine the value of V_e more accurately than by a first derivative curve.

The above equation simplifies to: $y = mx + b$ by equating

$$y = (V_a \cdot 10^{+\text{pH}}), m \text{ (the slope)} = -\left(\frac{1}{K_a} \frac{\gamma_B}{\gamma_{\text{BH}^+}} \right); x = (V_a) \text{ and } b = -\left(\frac{1}{K_a} \frac{\gamma_B}{\gamma_{\text{BH}^+}} \right) V_e.$$

The x-intercept occurs when $y=0$. That occurs when $V_a = V_e$. It is not necessary to know the value of the slope unless you want to evaluate the K_a value. In this case, one can

approximate that in the linear region of the curve, $\frac{\gamma_B}{\gamma_{\text{BH}^+}} \approx 1$.

***Submit a copy of your graph with the informal results for this experiment.

Waste Disposal: Discard all waste in appropriate bottles as instructed. If you have any questions, please see your instructor or teaching assistant before disposing of waste.

Informal report:

Experiment: Determination of Soda Ash

date submitted: _____

time submitted: _____

Name: _____

Values obtained:

(1) _____ %

Locker # _____

(indicator method)

(2) _____ %

(Gran plot)

(3) _____ %

Average value = _____ %

(if applicable) Standard deviation= _____ %

	I (indicator)	II(indicator)	III(GranPlot)
g dried Na ₂ CO ₃ used for standard solution:	_____	_____	_____ g

Standardization titration:

mLs of HCl added to reach 1 st end point	_____	_____	_____ mL
---	-------	-------	----------

mLs of HCl added to reach 2 nd end point	_____	_____	_____ mL
---	-------	-------	----------

Molarity of the HCl standard:	_____ M		
-------------------------------	---------	--	--

Sample calculation: M_{HCl} =

Unknown titration:

	I (indicator)	II(indicator)	III(GranPlot)
g dried unknown sample:	_____	_____	_____ g

mLs of HCl added to reach 1 st end point	_____	_____	_____ mL
---	-------	-------	----------

mLs of HCl added to reach 2 nd end point	_____	_____	_____ mL
---	-------	-------	----------

% Na ₂ CO ₃ in unknown: %	_____	_____	_____
---	-------	-------	-------

Sample calculation: Use the 2nd equivalence point: milliequivalents, mEq of titrant, A = mEq analyte, B at 2nd end point: mEq HCl = mEq Na₂CO₃; #mols HCl = 2 x(#mols Na₂CO₃);

$$M_{HCl} V_{HCl} = \frac{2(g_{Na_2CO_3})}{FW_{Na_2CO_3}} \Rightarrow g_{Na_2CO_3} = \frac{M_{HCl} V_{HCl} FW_{Na_2CO_3}}{2}$$

$$\% Na_2CO_3 = \frac{M_{HCl} V_{HCl} FW_{Na_2CO_3} (100\%)}{2(g_{ore})}$$

FW of Na₂CO₃ = 105.99 g/mol; FW of Na₂O = 61.98 g/mol;